



HPT Sensor Calibration Certificate

Material Number: 210091

Maximum Pressure (psi): 100

Serial Number: XD31353A

Calibration date: May 28, 2020

Technician: RSH

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.982E-1	-1.614E+0

Calibration PASSED