



## HPT Sensor Calibration Certificate

Material Number: 228265

Maximum Pressure (psi): 100

Serial Number: F3940H

Calibration date: April 01, 2021

Technician: RSH

### Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	2.577E+0	-2.551E+1

### Calibration FAILED

Test	Applied	Measured	Calculated	Diff.
ATM	14.740	15.723	15.008	-0.268
20 PSI	19.350	17.179	18.760	0.590
30 PSI	30.550	21.834	30.756	-0.206
40 PSI	40.420	25.645	40.577	-0.157
50 PSI	50.440	29.520	50.563	-0.123
60 PSI	60.030	33.212	60.077	-0.047
70 PSI	60.230	33.245	60.162	0.068
80 PSI	77.670	39.989	77.542	0.128