



HPT Sensor Calibration Certificate

Material Number: 228265
 Maximum Pressure (psi): 100

Serial Number: F3950H
 Calibration date: April 01, 2021
 Technician: RSH

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	2.584E+0	-2.613E+1

Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.740	15.745	14.555	0.185
20 PSI	20.620	18.085	20.602	0.018
30 PSI	30.340	21.984	30.677	-0.337
40 PSI	40.410	25.735	40.369	0.041
50 PSI	50.300	29.572	50.284	0.016
60 PSI	60.120	33.382	60.129	-0.009
70 PSI	70.070	37.217	70.039	0.031
80 PSI	79.790	40.990	79.788	0.002