



HPT Sensor Calibration Certificate

Material Number: 228265

Maximum Pressure (psi): 100

Serial Number: F396OH

Calibration date: April 01, 2021

Technician: RSH

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	2.575E+0	-2.575E+1

Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.740	15.724	14.739	0.001
20 PSI	20.740	18.060	20.754	-0.014
30 PSI	30.670	21.894	30.627	0.043
40 PSI	40.470	25.730	40.505	-0.035
50 PSI	50.280	29.534	50.300	-0.020
60 PSI	60.310	33.423	60.314	-0.004
70 PSI	70.140	37.288	70.267	-0.127
80 PSI	79.850	40.978	79.768	0.082