



HPT Sensor Calibration Certificate

Part Number: 210091
 Maximum Pressure (psi): 100

Serial Number: XD31087A
 Calibration date: July 10, 2018
 Technician: SPB

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.944E-1	-1.379E+0

Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.680	16.128	14.659	0.021
20 PSI	21.090	22.592	21.086	0.004
30 PSI	31.330	32.893	31.330	0.000
40 PSI	40.900	42.536	40.919	-0.019
50 PSI	50.980	52.675	51.001	-0.021
60 PSI	61.040	62.800	61.069	-0.029
70 PSI	70.620	72.410	70.626	-0.006
80 PSI	80.670	82.476	80.635	0.035