



## HPT Sensor Calibration Certificate

Part Number: 210091  
 Maximum Pressure (psi): 100

Serial Number: XD31088A  
 Calibration date: July 10, 2018  
 Technician: SPB

### Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.942E-1	-1.207E+0

### Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.680	15.950	14.650	0.030
20 PSI	20.380	21.693	20.360	0.020
30 PSI	29.980	31.382	29.993	-0.013
40 PSI	40.630	42.105	40.654	-0.024
50 PSI	50.550	52.088	50.579	-0.029
60 PSI	60.650	62.239	60.671	-0.021
70 PSI	70.840	72.462	70.835	0.005
80 PSI	80.280	81.916	80.234	0.046