



## HPT Sensor Calibration Certificate

Part Number: 210091

Maximum Pressure (psi): 100

Serial Number: XD20377A

Calibration date: September 29, 2014

Technician: SPB

### Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.888E-1	-1.356E+0

Calibration PASSED

Test	Applied	Measured
ATM	14.630	16.145
20 PSI	20.900	22.473
30 PSI	29.790	31.428
40 PSI	39.980	41.669
50 PSI	49.840	51.581
60 PSI	59.770	61.548
70 PSI	69.800	71.589
80 PSI	80.150	81.949