



## HPT Sensor Calibration Certificate

Part Number: 210091  
 Maximum Pressure (psi): 100

Serial Number: XD31039A  
 Calibration date: May 16, 2018  
 Technician: SD

### Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.956E-1	-1.122E+0

### Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.660	15.815	14.623	0.037
20 PSI	20.160	21.361	20.145	0.015
30 PSI	30.360	31.633	30.372	-0.012
40 PSI	40.220	41.558	40.253	-0.033
50 PSI	50.130	51.511	50.162	-0.032
60 PSI	60.300	61.721	60.327	-0.027
70 PSI	70.320	71.757	70.319	0.001
80 PSI	80.590	82.024	80.541	0.049