



HPT Sensor Calibration Certificate

Part Number: 210091
 Maximum Pressure (psi): 100

Serial Number: XD31040A
 Calibration date: May 16, 2018
 Technician: SD

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.965E-1	-1.075E+0

Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.630	15.726	14.596	0.034
20 PSI	20.150	21.276	20.127	0.023
30 PSI	30.210	31.408	30.223	-0.013
40 PSI	40.070	41.317	40.097	-0.027
50 PSI	50.060	51.347	50.092	-0.032
60 PSI	60.260	61.581	60.290	-0.030
70 PSI	70.230	71.555	70.230	0.000
80 PSI	79.960	81.262	79.903	0.057