



HPT Sensor Calibration Certificate

Part Number: 210091
 Maximum Pressure (psi): 100

Serial Number: XD31041A
 Calibration date: May 16, 2018
 Technician: SD

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.962E-1	-1.080E+0

Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.700	15.806	14.666	0.034
20 PSI	20.460	21.600	20.438	0.022
30 PSI	30.080	31.290	30.091	-0.011
40 PSI	40.180	41.442	40.205	-0.025
50 PSI	50.290	51.601	50.325	-0.035
60 PSI	60.510	61.852	60.537	-0.027
70 PSI	70.270	71.622	70.270	0.000
80 PSI	80.380	81.712	80.321	0.059