



## HPT Sensor Calibration Certificate

Part Number: 210091

Maximum Pressure (psi): 100

Serial Number: XD31119A

Calibration date: October 04, 2018

Technician: SPB

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.982E-1	-1.481E+0

Calibration PASSED