



HPT Sensor Calibration Certificate

Part Number: 210091

Maximum Pressure (psi): 100

Serial Number: XD31129A

Calibration date: December 11, 2018

Technician: SPB

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.966E-1	-1.554E+0

Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.600	16.157	14.548	0.052
20 PSI	20.060	21.660	20.032	0.028
30 PSI	30.300	31.977	30.314	-0.014
40 PSI	40.210	41.951	40.254	-0.044
50 PSI	50.190	51.970	50.239	-0.049
60 PSI	59.460	61.260	59.498	-0.038
70 PSI	70.330	72.129	70.330	0.000
80 PSI	80.630	82.387	80.553	0.077