



HPT Sensor Calibration Certificate

Part Number: 210091

Maximum Pressure (psi): 100

Serial Number: XD31132A

Calibration date: December 11, 2018

Technician: SPB

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.946E-1	-1.123E+0

Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.600	15.770	14.562	0.038
20 PSI	20.040	21.262	20.024	0.016
30 PSI	29.890	31.195	29.904	-0.014
40 PSI	41.690	43.080	41.724	-0.034
50 PSI	50.690	52.128	50.724	-0.034
60 PSI	60.330	61.818	60.361	-0.031
70 PSI	70.150	71.652	70.142	0.008
80 PSI	80.820	82.339	80.771	0.049