



HPT Sensor Calibration Certificate

Part Number: 210091

Maximum Pressure (psi): 100

Serial Number: XD31133A

Calibration date: December 11, 2018

Technician: SPB

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.975E-1	-1.210E+0

Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.600	15.804	14.554	0.046
20 PSI	20.230	21.470	20.206	0.024
30 PSI	30.250	31.556	30.267	-0.017
40 PSI	40.290	41.645	40.331	-0.041
50 PSI	50.380	51.771	50.432	-0.052
60 PSI	61.030	62.431	61.065	-0.035
70 PSI	71.110	72.495	71.104	0.006
80 PSI	80.950	82.308	80.892	0.058