



HPT Sensor Calibration Certificate

Part Number: 210091

Maximum Pressure (psi): 100

Serial Number: XD31145A

Calibration date: December 11, 2018

Technician: SPB

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.947E-1	-1.319E+0

Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.600	15.958	14.554	0.046
20 PSI	21.260	22.682	21.243	0.017
30 PSI	30.080	31.576	30.090	-0.010
40 PSI	40.330	41.908	40.367	-0.037
50 PSI	49.890	51.530	49.938	-0.048
60 PSI	60.500	62.176	60.527	-0.027
70 PSI	70.670	72.363	70.660	0.010
80 PSI	81.340	83.045	81.286	0.054