



## HPT Sensor Calibration Certificate

Part Number: 210091

Maximum Pressure (psi): 100

Serial Number: XD31157A

Calibration date: December 11, 2018

Technician: SPB

### Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.969E-1	-1.540E+0

### Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.600	16.138	14.548	0.052
20 PSI	21.890	23.489	21.876	0.014
30 PSI	30.710	32.358	30.718	-0.008
40 PSI	40.780	42.495	40.823	-0.043
50 PSI	51.040	52.786	51.082	-0.042
60 PSI	61.170	62.943	61.208	-0.038
70 PSI	71.150	72.926	71.160	-0.010
80 PSI	81.290	83.011	81.214	0.076