



## HPT Sensor Calibration Certificate

Part Number: 210091

Maximum Pressure (psi): 100

Serial Number: XD31159A

Calibration date: December 11, 2018

Technician: SPB

### Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.965E-1	-1.533E+0

### Calibration PASSED

Test	Applied	Measured	Calculated	Diff.
ATM	14.600	16.135	14.546	0.054
20 PSI	23.300	24.905	23.285	0.015
30 PSI	30.090	31.743	30.099	-0.009
40 PSI	39.950	41.662	39.983	-0.033
50 PSI	50.640	52.407	50.691	-0.051
60 PSI	60.670	62.455	60.703	-0.033
70 PSI	70.200	71.994	70.209	-0.009
80 PSI	80.560	82.301	80.480	0.080