



HPT Sensor Calibration Certificate

Material Number: 210091

Maximum Pressure (psi): 100

Serial Number: XD31222A

Calibration date: August 27, 2019

Technician: SPB

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.958E-1	-1.494E+0

Calibration FAILED

Test	Applied	Measured	Calculated	Diff.
ATM	14.610	16.124	14.562	0.048
20 PSI	19.930	21.486	19.902	0.028
30 PSI	30.200	31.849	30.221	-0.021
40 PSI	40.280	41.991	40.321	-0.041
50 PSI	50.180	51.941	50.229	-0.049
60 PSI	60.090	61.882	60.128	-0.038
70 PSI	70.220	72.017	70.221	-0.001
80 PSI	79.920	81.688	79.851	0.069