



HPT Sensor Calibration Certificate

Material Number: 210091

Maximum Pressure (psi): 100

Serial Number: XD31264A

Calibration date: November 08, 2019

Technician: SPB

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	9.963E-1	-1.303E+0

Calibration PASSED