



HPT Sensor Calibration Certificate

Material Number: 210091

Maximum Pressure (psi): 100

Serial Number: XD31297A

Calibration date: February 13, 2020

Technician: SPB

Calibration Values:

x5	x4	x3	x2	x1	x0
0	0	0	0	1.010E+0	-1.730E+0

Calibration PASSED